



**LAMY**  
**RHEOLOGY**  
INSTRUMENTS

# USING MANUAL B-ONE PLUS

VERSION N° B1-UK04/2021



## CONTENTS

<b>1</b>	<b>INTRODUCTION</b>	<b>3</b>
1.1	COMPONENTS	4
1.2	GENERAL VIEW OF YOUR DEVICE	5
1.3	CONNEXIONS	7
1.4	SPECIFICATIONS	7
1.5	INSTALLATION	8
<b>2</b>	<b>GETTING STARTED</b>	<b>9</b>
2.1	STATE ICONS	9
2.2	PRIMARY CONCEPTS	10
2.2.1	MAIN MENU	10
2.2.2	MEASURE	10
2.2.3	VIEW, EXPORT OR DELETE RESULT	12
2.2.4	ZERO SETTING	15
2.2.5	PARAMETERS	16
2.2.5.1	LANGUAGES	16
2.2.5.2	DATE / HOUR	17
2.2.5.3	SOUNDS / STANDBY / LIGHTING	17
2.2.5.4	USER NAME	18
2.2.5.5	UNITS	19
2.2.5.6	MEASURING SYSTEM	19
2.2.5.7	LOCKED MODE	21
2.2.5.8	TORQUE RANGE	22
2.2.5.9	MISCELLANEOUS	23
2.2.5.10	DENSITY	23
2.2.5.11	SERVICE	23
<b>3</b>	<b>MEASURING WITH YOUR DEVICE</b>	<b>24</b>
3.1	INSTALLATION OF MEASURING SYSTEM	24
3.2	MS RV/LV	24
3.3	MS BV	26
3.4	MS VANE	28
3.5	MS KREBS	29
3.6	MS DIN	30
3.7	MS SV (THERMOCELL and Small Volume Package)	36
3.8	MS ULV	41
3.9	T-BARS and HELIPRO	45
<b>4</b>	<b>VERIFICATION OF YOUR DEVICE</b>	<b>47</b>

## 1. INTRODUCTION

The B-ONE PLUS is a device able to measure the viscosity, which is the capacity of a product to resist to the flow.

The fluid is forced to a shear rate (rotational speed) and the shear stress (motor torque) is measured. The values of shear rate and shear stress then make it possible to calculate the viscosity using the Newton equation and the constants associated with the mobile used.

$$\text{Equation of Newton is: } \eta = \frac{\tau}{\dot{\gamma}}$$

With  $\eta$  for viscosity in Pa.s,  $\tau$  for shear stress in Pa and  $\dot{\gamma}$  for shear rate in  $s^{-1}$ .

Shear stress and shear rate are calculated by using constants of each measuring system as:

$$\tau = M \times K_{\tau} \text{ with } M \text{ for motor torque in mNm and } K_{\tau} \text{ in Pa/mNm.}$$

$$\dot{\gamma} = n \times K_D \text{ with } n \text{ for rotational speed in rpm and } K_D \text{ in } s^{-1}/\text{rpm.}$$

The viscometer calculates the viscosity by dividing the shear stress by the shear rate for each measuring point. The  $K_{\tau}$  and  $K_D$  constants used depend on the measuring system selected for the measurement.

Viscosity depends on the temperature, then it must be essential that all viscosity values are associated to a reading of the sample temperature, in order to compare viscosity for different samples.

There are some products for which the viscosity, to a constant temperature, stay unchanged, even if we change the shear rate. Those samples are named **Newtonian fluids**, i.e. : Oils, Water, Glycerol, etc...However, many substances have a variation of viscosity in function of speed of shearing, and the Flow Behaviour of those samples could be determined with measuring instruments able to set many speeds of rotation.

The viscometer is constituted with a continuous current motor with an optical encoder, in order to warranty a great accuracy of the speed of rotation of bob, on all torque range.

The viscometer has an easy touch screen display, on which you could read the **speed, measuring spindle** reference, temperature (in option), the measured torque and the dynamic **viscosity in mPa.s (=cPoises) or Pa.s**.

The viscometer B-ONE PLUS can be used with different measuring system. You will find below a list of compatible measuring system with this viscometer.

- **MS RV/LV:** Measuring spindles according to ASTM / ISO 2555 (316L stainless steel). These systems are ideally suited for simple viscosity measurement at controlled rotational speed in all areas of activity. The standard recommends use of 600ml beaker for measurement.

- **MS BV:** Measuring spindle for 150ml beaker (316L stainless steel). These spindles are ideally suited for simple viscosity measurement at a rotating speed in control in all areas of activity. They are appreciated for their ease of use and the low volume of product needed compared to the MS RV/LV spindles.

- **MS VANE:** Measuring spindles with blades (316L stainless steel). These systems are ideal for viscosity measurement (value or curve) in control or development of all types of products even of very high viscosity with or without particles (size <5mm). They can be used for direct measurement in user's containers or in tubes of MS-DIN systems.

- **MS KREBS:** Krebs type measuring spindles compatible with ASTM D562 standard (316L stainless steel). These systems are ideal for viscosity measurement in Krebs units in control of all types of products. They can be used for direct measurement in user containers or in 600 or 150ml beakers.

- **MS DIN:** Coaxial cylinders measuring systems according to DIN / ISO 3219 (316L stainless steel). These systems make it possible to set the shear rate in order to carry out viscosity measurements or to obtain curves to study flow behaviour, yield stress or thixotropy. They are particularly suitable for the control or development of homogeneous products with liquid aspect and with or without particles (size <200µm).


- **MS SV:** Measuring systems for low volumes (316L stainless steel). These systems, unlike the MS-ASTM and MS-DIN systems, make it possible to measure products in small quantities by applying a shear rate up to temperatures of 200 ° C (according to models, see table). With RT1, these systems are compatible with ASTM D3236 or D4402.

- **MS ULV:** Measuring system (316L stainless steel) for low viscosities usable with instruments LR version. This system, unlike the MS-RV/LV or MS-DIN systems, makes it possible to measure low viscosity products in control by applying a shear rate.

- **T-BARS:** Special spindle (316L stainless steel) used with HELIPRO stand. These spindle allow while move up/down of lift to avoid cavity formation. They are highly recommended with non-flowing materials.

### 1.1. COMPONENTS


Viscometer is delivered inside a foam protection to avoid any problem during transport. According to your order this foam can be inside a box or a carrying case and can contain measuring system.



In detail, you will find different part in your box as shown below.

Viscometer with rack stand.


Viscometer with standard stand



1. Anodized aluminium support.
2. Touch screen stylus.
3. Guide for Touch screen stylus.
4. Tool and screw.
5. Delrin safety ring.
6. Handle for arm.
7. Stop ring.
8. Notched rod and arm.
9. Smooth rod and arm.

## 1.2. GENERAL VIEW OF YOUR DEVICE

Once your device will be mounted and installed, it looks like this;




- TOUCH Screen

The new PLUS series is equipped with a 7" colour touch screen. It gives you greater working comfort and a clearer view of your data and analysis results.

- On / Off Switch

Always with the aim of improving your experience, LAMY RHEOLOGY has decided to equip all of its PLUS range with a luminous and design switch. It has been placed in the centre of the device for greater intuitiveness.




- Aluminium arm

The aluminium arm of the standard stand has tightening knob that allows you to maintain the height of the measuring head and handle for easy handling.

The aluminium arm of the rack support is fitted with a button for raising or lowering the measuring head.

The measuring head is attached to arm using a screw fitted with a tightening knob.




- **Stainless steel rod**

The rod of supports is made of stainless steel for a solid hold on the measuring head. They have a very long lifespan.

The standard support is fitted with a Delrin safety ring preventing the measuring head from touching the support. It is also equipped with a stop ring for memorizing a measurement height.

These two rings are not present on the rack support.




- **Anodized aluminium support**

The stand is entirely made of anodized aluminum. It gives our instruments an unmatched stability (the maximum permissible temperature on the white part is 50 ° C).

### 1.3. CONNEXIONS

According to your order, rear panel of device get this available connexions.



### 1.4. SPECIFICATIONS

**Type of instrument:** Rotating springless viscometer with 7" Touch screen

**Rotation speeds:** Unlimited number of speeds between 0.3 and 250 rpm

**Torque range:** Standard Version: 0.05 to 13 mNm. LR Version from 0.005 to 0.8mNm

**Accuracy:** +/- 1 % of the full scale

**Repeatability:** +/- 0,2 %

**Display:** Viscosity – Speed – Torque – Time - (Temperature in option) - Choice of viscosity units: cP/Poises or mPa.s / Pa.s

**Language:** French/English/Russian/Spanish/Turkish/German

**Compatible measuring system:** MS DIN, MS-KREBS, MS RV/LV, MS BV, MS VANE, MS ULV, MS SV, T-Bars.

**Compatible temperature control:** EVA LR-BV, RT-1, EVA MS DIN

**Supply voltage:** 90-240 VAC 50/60 Hz

**Connection:** USB

**Options:** See brochure

**Dimensions and weight:** Head: L180 x W130 x H230 mm, Hardened steel stand: L280 x W200 x H30 mm, Stainless steel rod: Length 500 mm, Weight: 6.7 kg

## 1.5. INSTALLATION


Your viscometer should be installed in a clean, vibration-free environment. Even if no level is necessary, choose a stable and flat table.

This installation notice concern only viscometer without any temperature device. If you have it, please refer to specific notice of temperature unit for installation.

For standard and rack stand: After unpacking all accessories from the box (or case if it is supplied), you must first screw the rod to the base with the help of the screw and the key provided.



For standard stand: Then make sure to insert the white Delrin cylinder and set it to the lowest position. Then place the stop ring on the stainless steel rod. Then place arm on rod, respecting the direction as described below. The arm is equipped with two buttons and a handle. You can change them for your convenience if you are left-handed.



For standard and rack stand: You can then put the viscometer on the arm, taking care not to touch the motor shaft or temperature sensor (if your device has one). Align the measuring head correctly and secure it with the screw provided.



Connect your viscometer by plugging power cable on to rear panel of device.



Your viscometer will be used with different measuring system. To know how to mount and use it, please refer section 3.

## 2. GETTING STARTED

Once power cable has been plugged on rear panel of device (see section 1.3), you can click on button to switch on your device (see section 1.2).

### 2.1. STATE ICONS

Once your device is switched on, you will see some icons on Touch Screen.


	No USB port used
	Port USB used
	Temperature reading of external Pt100 probe
	Enable to go to parameters of instrument.
	Enable to come back to Main Menu.
	Enable to come back to previous menu.

Some B-ONE PLUS are equipped with an external temperature sensor. When it has been ordered, the temperature display is shown at the top right of the screen. If this is not the case, see section 2.2.5.9 to enable reading. The USB port icon indicates that a USB flash drive has been connected. This is used to update the firmware or to transfer the results.

## 2.2. PRIMARY CONCEPTS

### 2.2.1. MAIN MENU

Main menu enable to you to browse between different tabs of your B-ONE PLUS. Acces is always available by clicking .



Button "Zero setting" is not present on this screen when you are using Version LR. Please see section 2.2.2. to see how to find similar button on LR device.


### 2.2.2. MEASURE

Measure tab is central part of your B-ONE PLUS. Before to use it, you should install your measuring system and your sample. Please see section 3.

Then you click on "Measure", you will see a new window.



Manual Mode enables to choose your measurement parameters like «Measuring System», «Speed or shear rate» then «Time of measurement».



**Rq :** If « Time » = 0, you could modify « speed » during the measurement. This could help you to define the best conditions to work on your sample.

If your measuring system is not in list, you may have to create it. Please refer to section 2.2.5.6.

Choice between “Speed” or “Shear rate” is according to your measuring system. For MS RV/LV, MS BV, MS KREBS and MS VANE, you should have only possibility to set the speed. For all other measuring system, you can use shear rate or speed if you force it to show rpm (see section 2.2.5.9). If you need to know what is the corresponding speed then you are using shear rate, you have to use constant  $K_D$  of your measuring system (information available in section 2.2.5.6).

$$\text{SPEED} = \text{SHEAR RATE} / K_D$$

With speed unit in rpm, shear rate in  $s^{-1}$  and  $K_D$  is  $rpm/s^{-1}$ .

Then your settings are ok, you can click “Start” to start your measurement.

With LR version, device will ask you to make zero of motor before measurement. Soon you have done it, click on back arrow and follow next step as standard version.

During your measurement, you will see a torque gage (on the bottom side of the display). Boundaries of this gage give you minimum and maximum viscosity you can measure with your selected spindle and set speed/shear rate. You have also value in % corresponding of measured torque vs maximum torque of device. This maximum torque or viewing % can be set on device (see section 2.2.5.8 and 2.2.5.9).



You must verify that the measured torque is not too close to the upper or lower limit, because you can get message as “Lower Torque” or “Torque Overload” and measurement will stop automatically. If this is the case, increase speed/shear rate or take a larger measurement system if you are close to the lower limit. Please decrease speed/shear rate or choose a smaller measurement system if the torque reading is close to the upper limit.


You will find several information available on the screen such as torque (mN.m), stress (Pa), temperature (° C), time (s) or viscosity (mPa.s). If the units do not suit you, you can change them in parameters (see section 2.2.5.5).

Then your measurement is finished, you will get this windows below. You will find all data you need and get possibility to save them into internal memory. If you choose “Save”, viscometer will ask you to give a name of your measurement. You will have after possibility to read it later (see section 2.2.3.).



### 2.2.3. VIEW, EXPORT OR DELETE RESULT

This menu allow you to read, export or delete data from internal memory. Press on « View results » tab in Main menu.



You will have two choice: See list of data to read them or delete data.

### 2.2.3.1. READ SAVED DATA

By click on tab “List of measurement data” you could see all saved measurement made with your B-ONE PLUS. You could select which one you want to read.



### 2.2.3.2. EXPORT DATA

Then USB flash drive is connected, “Global Export” will give you possibility to transfer all saved measure and “Export” will transfer only shown data.

The format of the data generated and saved by the viscometer is ASCII (\*.txt). Once your data has been copied to the USB drive, you can open the files using the EXCEL spreadsheet. To do this, simply copy the data from the USB key to your computer. Then open Excel, choose "File", "Open", taking care to select "All files \*. \*".



Find your file and click “Open”. Excel will offer you to convert your data by displaying three successive windows.


Select “Unicode UTF8” and “My data has Header” (last selection available according Office version) and click “Next”.



On second step, it will be necessary to choose the option “Tab” and “Semicolon” for separation of the columns. For “Text qualifier”, you should set “none”.



On last step, please choose “General” and click “Finish”.



You can then see your measurement results with the possibility to save a new file in Excel format.

### 2.2.3.3. DELETE SAVED DATA

By click on tab you could delete all saved measure one by one as you want from your B-ONE PLUS memory.




Then you click on “Delete”, saved data will be completely deleted from internal memory without any new confirmation.

## 2.2.4. ZERO SETTING

The zero setting allows you to calibrate your B-ONE PLUS to take account of the engine's empty friction. This function is not available on main menu for device in LR version. To find zero with LR device, please see section 2.2.2.




Standard Version



LR Version

The rotation speed for zero adjustment can be changed to suit your needs, giving you much more accurate measurements at specific speeds close to your measurement parameters.

For LR device, zero will be done at speed/shear rate set for manual measurement.



Standard Version



LR Version

This operation must be done without measuring system for standard device and with spindle for LR device. Then zero is finish you can click on OK and internal motor friction will be automatically saved inside memory of viscometer. If problem occur during zero setting, please try again. If problem still present, please contact your local distributor or society LAMY RHEOLOGY.

## 2.2.5. PARAMETERS

This parameters menu allow you to change settings of your device. It is reachable by clicking on icon “wheel” in upper left corner of touch screen.

This icon is only available then you are in “Main menu”.




### 2.2.5.1. LANGUAGES

Enable you to select language of your B-ONE PLUS. You have choice between French, English, Russian and Spanish. Then you have selected your desired language, you have to click on “Ok” and device will reboot automatically to show new language. In this menu you will be able to see Firmware version of your device.



### 2.2.5.2. DATE / HOUR


Enable you to adjust hour and date of your B-ONE PLUS. On this location, you can also set date for next checking of device. Soon this date will be reached, device will show you message as device need to be checked.



### 2.2.5.3. SOUNDS / STANDBY / LIGHTING

Allow you to modify sounds, lighting and activate or not the Standby mode of your B-ONE PLUS.

Choose if you want to get sound during using touch screen.



Choose if you want to get sound then measurement is finished.

Choose if you want to change brightness of Touch Screen.

Choose if you want to switch off automatically your device after no using. After selecting "Standby enable", you will have to set time. Your device will be switch off after this time.

#### 2.2.5.4. USER NAME

Operator mode will allow you to create different operators for your B-ONE. The use of the operators makes it possible to identify the person making the measurement (to save name on saved file and see later who made this measurement) or lock some function on device for simple user.


Operator management must always begin with the creation of the first account, which will become the administrator and thus create or delete another operator account. Click on "Create new user with PIN code". After specifying the name and password, the administrator will be named in red in the list.



You can now create another operator. The account of an operator may or may not be associated with a password (here called PIN code).

To delete an account, the administrator account must be used. Select the account you want to delete from the list and click on "Delete user name".

To use the operator accounts you must activate the mode by click "Enable User mode". Device will ask you to select user name you want to use. By returning to the Main Menu, you will see the name of the operator in use. By clicking on the arrow below the name of the operator, you can switch off the B-ONE or change operator. Click on "Logout" and device will ask you operator account you want to use




If the instrument is switched OFF and ON while operator mode is activated, device will ask you to select the operator you want use.

When User mode is enabled, some functions will be not editable for simple user as picture below shows it.




### 2.2.5.5. UNITS

Enable to you to change unit of viscosity values.



### 2.2.5.6. MEASURING SYSTEM

Allows you to add or remove a Measurement System.



All measurement systems stored by default in memory are not removable. Only those you have created yourself can be removed. To delete a measuring system, select it from the list and choose "Delete Measuring System". If this function remains greyed out when you have selected a system, it is part of the default mobile stored in the instrument's memory.

To add a new measuring system, you have to use "Ref of new system" function. Device will ask you name and constants for this measuring system.

You are not allowed to change the constant of an existing measuring system. If you want to use a new constant for an existing measuring system, you have to create a new one. Note that the KD constant is used to convert rotational speed to shear rate and K $\tau$  to convert torque to shear stress. Shear rate and shear stress are used to calculate the viscosity value. If you use a different constant value, you will get a different viscosity result.

Here is the list of constants used for measuring systems compatible with the B-ONE PLUS.

#### **MS BV**

<b>SYSTEM</b>	<b>K<math>\tau</math> / 1 mNm in Pa</b>	<b>Kd / 1 RPM in S-1</b>	<b>Ri / Ra</b>
<b>BV 1</b>	6,1	1,001	1
<b>BV 10</b>	25,5	0,5	0,7
<b>BV 100</b>	76,5	0,15	0,5
<b>BV 1000</b>	510	0,1	0,5

#### **MS RV/LV**

<b>SYSTEM</b>	<b>K<math>\tau</math> / 1 mNm in Pa</b>	<b>Kd / 1 RPM in S-1</b>	<b>Ri / Ra</b>
<b>RV 1</b>	<b>13,91</b>	<b>1</b>	<b>1</b>
<b>RV 2</b>	<b>55,65</b>	<b>1</b>	<b>1</b>
<b>RV 3</b>	<b>139,1</b>	<b>1</b>	<b>1</b>
<b>RV 4</b>	<b>278,2</b>	<b>1</b>	<b>1</b>
<b>RV 5</b>	<b>556,5</b>	<b>1</b>	<b>1</b>
<b>RV 6</b>	<b>1391</b>	<b>1</b>	<b>1</b>
<b>RV 7</b>	<b>5565</b>	<b>1</b>	<b>1</b>
<b>LV 1</b>	<b>106</b>	<b>1</b>	<b>1</b>
<b>LV 2</b>	<b>500</b>	<b>1</b>	<b>1</b>
<b>LV 3</b>	<b>1900</b>	<b>1</b>	<b>1</b>
<b>LV 4</b>	<b>8600</b>	<b>1</b>	<b>1</b>
<b>LV 5</b>	<b>17826</b>	<b>1</b>	<b>1</b>

#### **MS VANE**

<b>SYSTEM</b>	<b>K<math>\tau</math> / 1 mNm in Pa</b>	<b>Kd / 1 RPM in S-1</b>	<b>Ri / Ra</b>
<b>V71</b>	<b>36.5</b>	<b>1</b>	<b>0.5</b>
<b>V72</b>	<b>157</b>	<b>1</b>	<b>0.5</b>
<b>V72/2</b>	<b>270</b>	<b>1</b>	<b>0.5</b>
<b>V72/4</b>	<b>400</b>	<b>1</b>	<b>0.5</b>
<b>V72/6P</b>	<b>150</b>	<b>1</b>	<b>0.5</b>
<b>V-73</b>	<b>785</b>	<b>1</b>	<b>0.5</b>
<b>V-74</b>	<b>7850</b>	<b>1</b>	<b>0.5</b>
<b>V-75</b>	<b>2965</b>	<b>1</b>	<b>0.5</b>
<b>VT105</b>	<b>2180</b>	<b>1</b>	<b>0.5</b>
<b>VT2010</b>	<b>410</b>	<b>1</b>	<b>0.5</b>
<b>VT2020</b>	<b>59</b>	<b>1</b>	<b>0.5</b>
<b>VT3015</b>	<b>80</b>	<b>1</b>	<b>0.5</b>
<b>VT4020</b>	<b>34</b>	<b>1</b>	<b>0.5</b>
<b>VT4040</b>	<b>7.4</b>	<b>1</b>	<b>0.5</b>
<b>VT5025</b>	<b>17</b>	<b>1</b>	<b>0.5</b>
<b>VT6015</b>	<b>43</b>	<b>1</b>	<b>0.5</b>
<b>VT6030</b>	<b>10</b>	<b>1</b>	<b>0.5</b>
<b>VT608</b>	<b>150</b>	<b>1</b>	<b>0.5</b>
<b>VT8040</b>	<b>4.2</b>	<b>1</b>	<b>0.5</b>
<b>VT8070</b>	<b>1.2</b>	<b>1</b>	<b>0.5</b>

## MS DIN

SYSTEM	Ktau / 1 mNm in Pa	Kd / 1 RPM in S-1	Ri / Ra
MS-DIN 11	13.2	1.291	0.92
MS-DIN 12	19.4	0.354	0.73
MS-DIN 13	64.4	0.152	0.43
MS-DIN 22	25.8	1.291	0.92
MS-DIN 23	77.9	0.19	0.54
MS-DIN 33	130.1	1.291	0.92
MS-DIN 19	12.56	3.223	0.97

## MS SV and MS ULV

SYSTEM	Ktau / 1 mNm in Pa	Kd / 1 RPM in S-1	Ri / Ra
SV414	877	0.4	0.69
SV415	371	0.48	0.75
SV416	572	0.29	0.53
SV418	59.7	1.32	0.92
SV421	65.9	0.93	0.88
SV425	1918	0.22	0.25
SV427	126.8	0.34	0.62
SV428	205.2	0.28	0.49
SV429	367	0.25	0.40
SV431	166.5	0.338	0.62
SV434	271	0.28	0.49
SVC	68	0.43	0.71
SVTR8	66.15	0.92	0.88
SVTR9	127	0.34	0.62
SVTR10	204	0.28	0.49
SVTR11	374	0.25	0.40
MS-ULV	33.1	2.04	0.95

## T-Bars

SYSTEM	Ktau / 1 mNm in Pa	Kd / 1 RPM in S-1	Ri / Ra
T-A 92	278	1	1
T-B 93	557	1	1
T-C 94	1392	1	1
T-D 95	2783	1	1
T-E 96	6957	1	1
T-F 97	13914	1	1

### 2.2.5.7. LOCKED MODE

This option allow you to block measuring parameters. It should be set by an administrator or responsible of the device.

This function is not comparable to the "User name" menu (please see section 2.2.5.4). It should be use if you want to protect measurement settings on your device. All settings will be not lock by this function. You will see below which settings are concerned.

This function will block also parameters for measure. In this way, if you want to use all the time same parameters for measurement, you should enable this locked mode to be sure that nobody will change settings for measurement.



When you click "Enable", the B-ONE will ask you to save a 4-digit code that will be required to disable this protected mode. Each activation is independent and can be done with a different code. The protected mode is indicated by the presence of a padlock-like icon. **BUT TO DISABLE THIS LOCKED MODE, YOU SHOULD USE 4-digit CODE USED TO ENABLE IT.**

Once protected mode is activated, you will see this icon on B-ONE Screen (see picture below). Protected mode protect programs, measuring parameters and some menu as shown on pictures below.



### 2.2.5.8. TORQUE RANGE

Enable you to adjust the torque range of device according to spring viscometer technology. This setting will have effect on torque in % shown while measurement and viscosity limits.



Standard Version


LR Version

Please see below torque covered by specific range:

- LV Range : From 0.00673 to 0.0673 mNm.
- RV Range : From 0.07187 to 0.7187 mNm.
- HA Range : From 0.1437 to 1.4374 mNm.
- HB Range : From 0.5749 to 5.7496 mNm;
- None means no limits. So it will be complete range for standard device (from 0.05 to 13 mNm) and LR device (from 0.005 to 0.8 mNm).

### 2.2.5.9. MISCELLANEOUS

This menu allows you to select the temperature sensor used by the viscometer. The B-ONE allows the use of an external probe (Pt100-2) which must be connected to the back of the viscometer (see section 1.3). This setting is not available by default on B-ONE PLUS if this device is delivered without external temperature probe. If you buy it later and want to activate this function, please contact your local agent or LAMY RHEOLOGY.



“Update” function is used when updating the machine software is necessary. Do not go in this menu without being invited by the company LAMY RHEOLOGY. The update is done via a USB key connected to the "USB" port. You can then click on "Update" to update your instrument. At the end, your device will turn off and you will have to turn it on again.

“Operating time” and “Measuring time” settings indicate time while device has been switched ON and time while it was used for measurement.

“Forcing rpm” allow you to force device setting only speed in rpm instead shear rate (s-1) when you are using measuring system compliant with DIN 53019 norm (as MS DIN, MS ULV, MS SV).

“Enable torque % display” allow device to show torque in % above torque gage while measurement.

### 2.2.5.10. DENSITY

Enable you to enter density value of your product to measure in order to calculate his kinematic viscosity.



If you set a density value, you will get all the time viscosity in cStoke. Please remove density information if you want to get back Pa.s or Poise for unit of viscosity.

### 2.2.5.11. SERVICE

Reserved to LAMY RHEOLOGY engineers.

### 3. MEASURING WITH YOUR DEVICE


This section will show how use the different measuring system with your device.

Viscometer need to be installed before next section of this manual (see section 1.5).

#### 3.1. INSTALLATION OF MEASURING SYSTEM

Read the installation of your measuring system in the following sections before inserting it on your viscometer. Indeed some measuring systems require the installation of accessory before the insertion of the spindle.

As the B-ONE PLUS get only one kind of bayonet coupling system, way to install measuring bob on shaft of viscometer is always the same.



Insert the measuring system with the bayonet coupling into the motor shaft by pushing and turning slightly so that the pin is lodged in the space provided.

#### 3.2. MS RV/LV

Measuring spindles according to ASTM / ISO 2555 (316L stainless steel).

These systems are ideally suited for simple viscosity measurement at controlled rotational speed in all areas of activity. The standard recommends use of 600ml beaker for measurement.

Here below are all available mobiles:

Name	Part number	Dim. (mm)
LV-1 spindle	111010	Ø 18,80 - L 65,1
LV-2 spindle	111011	Ø 18,72 - L 6,86
LV-3 spindle	111012	Ø 12,60 - L 1,78
LV-4 spindle	111013	Ø 3,20 - L 31
Axis R 1-6 without disc	111000	Axe fileté
RV-1 Disc	111001	Ø 56,26
RV-2 Disc	111002	Ø 46,93
RV-3 Disc	111003	Ø 34,69
RV-4 Disc	111004	Ø 27,30
RV-5 Disc	111005	Ø 21,14
RV-6 Disc	111006	Ø 14,62
RV-7 Spindle	111007	Ø 3,20

The spindle L are delivered complete, while the R discs must be screwed on the R1-6 axis (Ref 111000).

These spindle are composed of two groups. The mobiles L are intended for low viscosity fluids and R mobiles for medium to high viscosities (see tables below):

Designation spindle	Part Number Spindle	Part Number Complete set**		B-ONE LR	B-ONE
RV1	111001*	111947	111948	Not Usable	100 to 0.6M
RV2	111002*			200 to 0.14M	200 to 2.4M
RV3	111003*			300 to 0.37M	300 to 6M
RV4	111004*			400 to 0.74M	600 to 12M
RV5	111005*			500 to 1.4M	1.2K to 24M
RV6	111006*			1200 to 3.7M	2.8K to 60M
RV7	111007			4500 to 15M	12K to 240M
LV1	111010	111014		15 to 0.25M	200 to 4.3M
LV2	111011			50 to 1.3M	1K to 20M
LV3	111012			200 to 5M	4k to 82M
LV4	111013			1000 to 22M	17K to 370M

M for millions, K for thousand

a) Need additional axis (PN111000)

b) Complete set (delivered with axis PN 111000 only for RV spindle)

When measuring, it is strongly recommended to heat the 600ml beaker. You can use either a thermostatic bath or the EVA LR PLUS temperature control system.

Place the viscometer on its support (see section 1.5). Fill the beaker with 500 ml of product to be tested, taking care not to introduce air bubbles.

Place it in a bath (if you have one) for a sufficient time to reach the desired temperature.

If the product contains volatile or hygroscopic material, cover the beaker for the duration of the operation.

Place the measuring head in the highest position (use screw on aluminium arm as wrote in section 1.2).

Make a zero of your viscometer (see section 2.2.4.) if you are using standard model.

Insert the measuring system with the bayonet coupling in the motor shaft (see section 3.1).

Make a zero of your viscometer (see section 2.2.4.) if you are using LR model.



Standard Stand



Rack Stand

Use handle to manipulate the device on standard stand (see section 1.2), use screw on aluminium arm and go down to immerse the mobile in the product. With device on Rack stand, just turn screw to move down the measuring head. Beware of air bubbles under the disk!

Adjust the position of the viscometer in the sample to immerse the mobile to the predefined mark (the lowest for discs # 2 -6, the highest for disc # 1), so as to immerse the Pt100 probe at least 3 mm (only for models equipped with a temperature probe integrated into the measuring head. If an external temperature sensor is used, the liquid level must always be in the mark on the axis mobile). Be careful that the tip of the mobile is 10 mm or more from the bottom of the beaker.





When the measurement height is optimum, use screw on arm to block height of measuring head when you are using standard stand. After you can use the stop ring on the support rod to memorize the position. With rack stand, you don't need to lock position and stop ring is not provided.

Wait until the temperature of the sample is within the prescribed limits (if you have temperature probe with your device).

Start the measurement at the desired speed and after choose the right measuring system (see section 2.2.2).

The torque measurement is indicated on the instrument screen using a gage. Ensure that the measured torque is always sufficiently far from the lower and upper limits (at least 5% above and below). If this is not the case, you can either change the measuring spindle or change the rotating speed.

Then your measurement is finished, raise the measuring head and lock it with screw on aluminium arm as wrote in section 1.2. Remove the measuring spindle to clean it.

### 3.3. [MS BV](#)

Measuring spindle for 150ml beaker (316L stainless steel).

These spindles are ideally suited for simple viscosity measurement at a rotating speed in control in all areas of activity. They are appreciated for their ease of use and the low volume of product needed compared to the MS RV/LV spindles.

Here below are all available spindle:

	Name	Part number	Dim. (mm)
	BV 1-100 Axis	117102	-
	BV centring device	117202	-
	BV Disc n°1	117001	Ø 45
	BV Disc n°10	117010	Ø 40
	BV Disc n°100	117100	Ø 20
	BV 1000 Axis	117101	Ø 4

This spindle allow measurement of huge viscosity range as showing below.

Designation spindle	Part Number Spindle <sup>b)</sup>	Part Number Complete set <sup>c)</sup>	Viscosity range for B-ONE (mPa.s)
BV1	117001 <sup>a)</sup>	117000	15 to 0.25M
BV10	117010 <sup>a)</sup>		100 to 2M
BV100	117100 <sup>a)</sup>		1K to 22M
BV1000	117101		10K to 220M

M for millions, K for thousand

Use specific glass Beaker (PN117150 for 10pcs) or specific plastic beaker (PN117155 for 10 pcs).

a) Need additional axis (PN 117102)

b) Need to be used with Centring piece (PN 117202)

c) Complete set delivered with axis (PN117102) and centring tool (PN 117202)

The BV 1000 Axis can be used like this. But for BV Disc 1, 10 and 100, you should screw it on BV 1-100 Axis.

When measuring, it is strongly recommended to heat the 150ml beaker. You can use either a thermostatic bath or the EVA BV PLUS temperature control system.

Place the viscometer on its support (see section 1.5). Fill the beaker with 120 ml of product to be tested, taking care not to introduce air bubbles.

Place it in a bath (if you have one) for a sufficient time to reach the desired temperature.

If the product contains volatile or hygroscopic material, cover the beaker for the duration of the operation.

Place the measuring head in the highest position (use screw on aluminium arm as wrote in section 1.2).

Make a zero of your viscometer (see section 2.2.4.) if you are using standard model.

Insert the measuring system with the bayonet coupling in the motor shaft (see section 3.1).

Make a zero of your viscometer (see section 2.2.4.) if you are using LR model.

Use handle to manipulate the device on standard stand (see picture section 3.2), use screw on aluminium arm and go down to immerse the mobile in the product. With device on Rack stand, just turn screw to move down the measuring head. Beware of air bubbles under the disk!



Adjust the position of the viscometer in the sample to immerse the mobile to the predefined mark. Be careful that the tip of the mobile is 10 mm or more from the bottom of the beaker.

When the measurement height is optimum, use screw on arm to block height of measuring head when you are using standard stand. After you can use the stop ring on the support rod to memorize the position. With rack stand, you don't need to lock position and stop ring is not provided (see picture on section 3.2).

Wait until the temperature of the sample is within the prescribed limits.

Start the measurement at the desired speed and after choose the right measuring system (see section 2.2.2).

The torque measurement is indicated on the instrument screen using a gage. Ensure that the measured torque is always sufficiently far from the lower and upper limits (at least 5% above and below). If this is not the case, you can either change the measuring spindle or change the measurement speed.

Then your measurement is finished, raise the measuring head and lock it with screw on aluminium arm as wrote in section 1.2. Remove the measuring spindle to clean it.

### 3.4. MS VANE

Measuring spindles with blades (316L stainless steel).

These systems are ideal for viscosity measurement (value or curve) in control or development of all types of products even of very high viscosity with or without particles (size <5mm). They can be used for direct measurement in user's containers.



All data given in table next page are given for information and can be changed according container use for measurement. For example, shear rate range show same data as for speed range of B-ONE PLUS. And most of the time, you will be able to use only speed for your viscosity measurement and not the shear rate.

Here below are all available spindles with viscosity range (in mPa.s):

Designation	Part Number	Diameter (mm)	Lenght (mm)	B-ONE LR	B-ONE
MK-V71	111114	34.39	68.78	1,4 to 18K	14 to 300K
MK-V72**	120017	21.67	43.38	5,6 to 74K	56 to 1,2M
MK-V73**	111108	12.67	25.35	28 to 370K	280 to 6M
MK-V74**	111115	5.89	11.76	280 to 3,7M	2,8K to 60M
MK-V75**	111111	8.026	16.05	111 to 1,4M	1,1K to 24M
MK-V72/2**	111112	21.67	20	54 to 720K	540 to 11M
MK-V72/4**	111113	21.67	10	80 to 1M	800 to 17M
MK-V72-6P*	111121	21.67	43	30 to 400K	300 to 6,5M
MK-VT105**	440105	5	10	430 to 5,8M	4,4K to 94M
MK-VT2010**	442010	10	20	82 to 1M	820 to 17M
MK-VT2020**	442020	20	20	12 to 150K	118 to 2,5M
MK-VT3015**	443015	15	30	16 to 210K	160 to 3,4M
MK-VT4020**	444020	20	40	7 to 90K	68 to 1,4M
MK-VT4040	444040	40	40	1,5 to 19K	15 to 320K
MK-VT5025**	445025	25	50	4 to 45K	34 to 730K
MK-VT6015	446015	15	60	9 to 114K	86 to 1,8M
MK-VT6030	446030	30	60	2 to 26K	20 to 433K
MK-VT608	446008	8	60	30 to 400K	300 to 6,5M
MK-VT8040	448040	40	80	1 to 11K	9 to 182K
MK-VT8070	448070	70	80	0,5 to 3,2K	3 to 52K

M for million, K for thousand

\* VANE 6 BLADES.

\*\* These items can be used with tube MB-DIN1 (P.N.112932).

All data given in this table are given for information and can be changed according container use for measurement. For example, shear rate range show same data as for speed range of B-ONE PLUS. And most of the time, you will be able to use only speed for your viscosity measurement and not the shear rate.

Place the viscometer on its support (see section 1.5).

Place the measuring head in the highest position (use screw on aluminium arm as wrote in section 1.2).

Make a zero of your viscometer (see section 2.2.4.) if you are using standard model.

Insert the measuring system with the bayonet coupling in the motor shaft (see section 3.1).

Make a zero of your viscometer (see section 2.2.4.) if you are using LR model.

For the use of these measuring system you can use 600ml or 150 ml beaker or your own container. Fill your beaker or container.

Use handle to manipulate the device on standard stand (see picture section 3.2), use screw on aluminium arm and go down to immerge the mobile in the product. With device on Rack stand, just turn screw to move down the measuring head.

Adjust the position of the viscometer to immerge the measuring system in the sample (only part with blade) and not to close from the bottom of beaker (minimum distance should be 10mm).

When the measurement height is optimum, use screw on arm to block height of measuring head when you are using standard stand. After you can use the stop ring on the support rod to memorize the position. With rack stand, you don't need to lock position and stop ring is not provided (see picture on section 3.2).

Start the measurement at the desired speed and after choose the right measuring system (see section 2.2.2). If your Vane measuring system is not in the list of device, please refer to section 2.2.5.6 to create it.

The torque measurement is indicated on the instrument screen using a gage. Ensure that the measured torque is always sufficiently far from the lower and upper limits (at least 5% above and below). If this is not the case, you can either change the measuring spindle or change the measurement speed.

Then your measurement is finished, raise the measuring head and lock it with screw on aluminium arm as wrote in section 1.2. Remove the measuring spindle to clean it.

### 3.5. MS KREBS

Krebs type measuring spindles compatible with ASTM D562 standard (316L stainless steel). These systems are ideal for viscosity measurement in Krebs units in control of all types of products. They can be used for direct measurement in user containers or in 600 or 150ml beakers.

Here below are all available spindle:

Name	Part number	Dim. (mm)	
MK-KU 1-10	111100	L. 54,11	
MK-75Y	111103	L. 34,58	

For your information, only MK-KU1-10 is conform to the norm ASTM D562.

To get KU unit for your viscosity measurement with your viscometer, you must choose the measuring spindle MK-KU1-10 and speed at 200 rpm. For all other speed and measuring spindle, you will get viscosity value in Pa.s.

Range for these spindle is (for LR on demand):

- MK KU1-10: 20-500mPa.s and 40-140KU (at 200 rpm).
- MK-75Y: 100-50000 mPa.s.

Place the viscometer on its support (see section 1.5).

Place the measuring head in the highest position (use screw on aluminium arm as wrote in section 1.2).

Make a zero of your viscometer (see section 2.2.4.) if you are using standard model.

Insert the measuring system with the bayonet coupling in the motor shaft (see section 3.1).

Make a zero of your viscometer (see section 2.2.4.) if you are using LR model.

For the use of these measuring system you can use 600ml or 150 ml beaker or your own container. Fill your beaker or container.

Use handle to manipulate the device on standard stand (see picture section 3.2), use screw on aluminium arm and go down to immerge the mobile in the product. With device on Rack stand, just turn screw to move down the measuring head.

Adjust the position of the viscometer to immerge the measuring system in the sample (only part with blade) and not to close from the bottom of beaker (minimum distance should be 10mm). Be careful that the tip of the mobile is 10 mm or more from the bottom of the beaker.

When the measurement height is optimum, use screw on arm to block height of measuring head when you are using standard stand. After you can use the stop ring on the support rod to memorize the position. With rack stand, you don't need to lock position and stop ring is not provided (see picture on section 3.2).

Start the measurement at the desired speed and after choose the right measuring system (see section 2.2.2). If your Vane measuring system is not in the list of device, please refer to section 2.2.5.6 to create it.

The torque measurement is indicated on the instrument screen using a gage. Ensure that the measured torque is always sufficiently far from the lower and upper limits (at least 5% above and below). If this is not the case, you can either change the measuring spindle or change the measurement speed.

Then your measurement is finished, raise the measuring head and lock it with screw on aluminium arm as wrote in section 1.2. Remove the measuring spindle to clean it.

### 3.6. MS DIN

Coaxial cylinders measuring systems according to DIN / ISO 3219 (316L stainless steel).

These systems make it possible to set the shear rate in order to carry out viscosity measurements or to obtain curves to study flow behaviour, yield stress or thixotropy.

They are particularly suitable for the control or development of homogeneous products with liquid aspect and with or without particles (size <200µm).

Please find below MS DIN items.

Name	Reference	
MK - DIN 1	112820	
MK - DIN 2	112821	
MK - DIN 3	112822	
MK - DIN 9	111875	

Name	Reference	
MB-DIN 1 Tube	112932	
MB-DIN 2 Tube	112937	
MB-DIN 3 Tube	112938	
CAP-DIN 1	112872	
CAP-DIN 2	112877	
CAP-DIN 3	112878	
CAP-DIN 1 Mooney	112874	
ST-R centring tool	114436	
MB-DIN 1 S Tube	112933	
MB-DIN 2 S Tube	112948	
MB-DIN 3 S Tube	112944	

Complete configurations include a DIN tube or DIN XS tube, a MK-DIN cylinder and a cap.

Measuring system		Spindle	Cup	Cap	Category
Designation	Part Number	Designation	Designation	Designation	
MS DIN 11	112801	MK-DIN1	MB-DIN1	CAP-DIN1	A
MS DIN 11 S	112809	MK-DIN1	MB-DIN1S	CAP-DIN1	B
MS DIN 12	112802	MK-DIN2	MB-DIN1	CAP-DIN1	A
MS DIN 13	112803	MK-DIN3	MB-DIN1	CAP-DIN1	A
MS DIN 13S	112808	MK-DIN3	MB-DIN1S	CAP-DIN1	B
MS DIN 19	112806	MK-DIN9	MB-DIN1	CAP-DIN1	A
MS DIN 22	112804	MK-DIN2	MB-DIN2	CAP-DIN2	A
MS DIN 22S	112815	MK-DIN2	MB-DIN2S	CAP-DIN2	B
MS DIN 33	112805	MK-DIN3	MB-DIN3	CAP-DIN3	A
MS DIN 33S	112814	MK-DIN3	MB-DIN3S	CAP-DIN3	B
MS DIN 11M	112812	MK-DIN1	MB-DIN1	CAP-MOONEY	A
MS DIN 19M	112811	MK-DIN9	MB-DIN1	CAP-MOONEY	A
MS DIN 23	112816	MK-DIN3	MB-DIN2	CAP-DIN2	A

Here are the measuring ranges (viscosity in mPa.s \*\*) of the existing MS DIN measuring systems:

Measuring system *		Volume (ml)	Shear rate (s <sup>-1</sup> )**	B-ONE LR	B-ONE
Designation	Part Number				
MS DIN 11	112801	27	1,29N	2,5 to 27K	25 to 0.44M
MS DIN 11S	112809***	27	1,29N	2,5 to 27K	25 to 0.44M
MS DIN 12	112802	46	0,35N	11 to 145K	110 to 2.3M
MS DIN 13	112803	61	0,15N	92 to 510K	920 to 8.3M
MS DIN 13S	112808***	22	0,15N	93 to 510K	920 to 8.3M
MS DIN 19	112806	25	3,22N	0,8 to 10K	8 to 0.17M
MS DIN 22	112804	22	1,29N	4 to 53K	40 to 0.86M
MS DIN 22S	112815***	22	1,29N	5 to 53K	40 to 0.86M
MS DIN 33	112805	14	1,29N	20 to 265K	200 to 4.3M
MS DIN 33S	112814***	14	1,29N	20 to 265K	200 to 4.3M
MS DIN 11M	112812	23	1,29N	2,5 to 27K	21 to 0.44M
MS DIN 19M	112811	18.5	3,22N	0,8 to 10K	8 to 0.17M
MS DIN 23	112816	36	0,19N	81 to 1M	810 to 17M

M for million, K for thousand, N for rotational speed (rpm)

\* Complete measuring system with spindle, cup and cap.

\*\* These values are given when complete system is used.

\*\*\* Include centring tool ST-R (P.N. 114436).

### 3.6.1. USE OF CATEGORY A

Each tube is used with the matching cylinder (eg DIN tube 1 with cylinder MK-DIN1). The cylinder MK-DIN 9 is used with the DIN tube 1. The tubes can be closed with their cap assorted or used open for a measurement immersed in a container containing the liquid to be measured. The Mooney plug is used exclusively with the DIN 1 tube and the MK-DIN 1 and MK-DIN 9 cylinders. It reduces the volume of product required for the measurement (see table on the next page).

In addition to these measuring systems, a DIN X tube with a MK-DIN X + 1 cylinder can be used. Thus, MS DIN 12, MS DIN 13 and MS DIN 23 measuring systems can also be used. The first digit always indicates the number of the tube and the second digit the number of the cylinder MK. There is also a measuring system MS DIN 19 which uses the DIN tube 1 and MK-DIN 9 cylinder.

All systems shown above can be used with or without a temperature control unit since the tube attaches to the base of the viscometer (see below). When your viscometer is combined with an EVA MS DIN, EVA MS DIN-MSR, EVA 100, RT1, CT DIN or CT-LC temperature setting unit, the positioning of the measuring cylinders and tubes is the same. It will be necessary to introduce the assembly into the temperature chamber.



The first step is to install the cap on the tube as shown in the picture below (not necessary if you make a dip measurement directly into a pot). Also check that the gasket is properly installed on the cap. The first insertion of the cap can be difficult. You must use some silicone grease to facilitate installation.



You can then put the product to be measured in the tube. The necessary volume is indicated in the table on the previous page according to the system used. There is a level line in the tube (see picture).

Place the viscometer on temperature unit or on its own stand if you are using CT-DIN or CT-LC. Place the measuring head in the highest position (use screw on aluminium arm as wrote in section 1.2).

Make a zero of your viscometer (see section 2.2.4.) if you are using standard model.

Insert the measuring system with the bayonet coupling in the motor shaft (see section 3.1).

Make a zero of your viscometer (see section 2.2.4.) if you are using LR model.

You can then introduce the MS DIN tube with the product from the bottom up slowly and making sure to align with the cylinder. Also place the groove on the tube facing you as shown in the picture below). When you are close enough to the viscometer basis, you must rotate the tube to place the pin in the groove of the tube.



Once the tube is properly installed, check that the temperature sensor (if you device gets one) on the viscometer is at least 3mm deep into the liquid.

When using your viscometer with a temperature control system (EVA MS DIN, EVA MS DIN MSR, EVA 100, CT DIN or RT-1), you must then lower the measuring head in such a way that the tube comes inside the hole provided for this purpose. Check that there is no product on the outer wall of the tube.



Use the arm handle and screw (see section 1.2) to gently lower the measuring head. The tube should be easily housed in the chamber. If this is not the case, do not force it and contact your local distributor or LAMY RHEOLOGY.

When the measuring head is properly positioned, there is no need to lock the position using the screw on the stem. Also, be sure to allow sufficient time for stabilization if the measurement temperature is different from the ambient temperature (at least 10 minutes) before making your measurement.

When using MS DIN measuring systems for immersion measurement, the order and method of installation is the same for cylinder and tube. Since there is no cap on the tube, you will need to put the product to be measured in a pot larger than the tube.



Then lower the measuring head by using the handle and the screw on the arm (see section 1.2) so that the tube is immersed in the pot and the temperature probe (if your device gets one) is sufficiently immersed in the liquid (at least 1 cm). Take care that the liquid level never exceeds the level of the white ring present on the cylinder axis MK DIN. You must use the screw on the arm (see section 1.2) to block the measuring position and take care that there is enough space between the tube and the bottom of the pot for the liquid to rise in the tube.

Once the setup is complete, you can do your measurement (see section 2.2.2).


If you have just made a measurement with the tube closed by a cap, rise the measuring head to the highest position by blocking the arm with the screw provided for this purpose (see section 1.2). Carefully remove the cylinder from the viscometer shaft, then remove the tube containing the product and the cylinder, being careful if the temperature is high. You can then remove the MK DIN cylinder from the DIN tube for cleaning. Remove the DIN tube cap to clean it. Remember to clean the temperature probe (if present).

If you have just made a measurement in immersion (without cap), it is preferable to remove the cylinder from the axis of the motor and after the tube from the base of the viscometer before rise the measuring head. You can then take the tube and cylinder for cleaning. Remember to clean the temperature probe (if present).

### 3.6.2. USE OF CATEGORY B

There are also other DIN x S type tubes that are used with EVA MS DIN, EVA MS DIN-MSR, EVA 100, RT1, CT DIN or CT-LC temperature units.

DIN xS tubes are used with the same MK-DIN cylinders and caps as standard DIN tubes. The measuring ranges are therefore unchanged (see table section 3.6). Their using facilitates cleaning and filling since they are shorter and therefore easier to access. On the other hand, they require the use of an accessory (Ref 114436) which guarantees a perfect centering of the cylinder in the tube.



The first step is to install the cap on the tube as shown in the photo below. Also check that the gasket is properly installed on the cap. The first insertion of the cap can be difficult. You must use some silicone grease to facilitate installation.



You can then put the product to be measured in the cylinder. The necessary volume is indicated in the table section 3.6 according to the system used. There is a level line in the tube (see picture below).

Place the viscometer on temperature unit or on its own stand if you are using CT-DIN or CT-LC. Place the measuring head in the highest position (use screw on aluminium arm as wrote in section 1.2).



You must then install the centring piece on the base of the viscometer.

Also place the groove facing you as shown in the picture below. When you are close enough to the viscometer basis, you must rotate the piece to place the pin in the groove.

Make a zero of your viscometer (see section 2.2.4.) if you are using standard model.

Insert the measuring system with the bayonet coupling in the motor shaft (see section 3.1).

Make a zero of your viscometer (see section 2.2.4.) if you are using LR model.

Put the tube with your product inside the warming unit.



Use the arm handle and screw (see section 1.2) to gently lower the measuring head. The centring piece must completely cover the rim of the tube.

When the measuring head is properly positioned, there is no need to lock the position using the screw on the arm. Also, be sure to allow sufficient time for stabilization if the measurement temperature is different from the ambient temperature (at least 10 minutes) before making your measurement.

Once the setup is complete, you can do your measurement (see section 2.2.2).

When your measurement is complete, it is better to remove the cylinder from the viscometer shaft. That will rest in the tube. Rise the measuring head to the highest position by blocking the arm with the screw provided (see section 1.2). Then remove the tube containing the product and the cylinder, being careful if the temperature is high. You can then remove the MK DIN cylinder from the tube for cleaning. Remove the DIN tube cap to clean it. Don't forget to clean the temperature probe (if available).

### 3.7. MS SV

Measuring systems for low volumes (316L stainless steel).

This section explain also how to use THERMOCELL and Small volume package.

These systems, unlike the MS-RV/LV and MS-DIN systems, make it possible to measure products in small quantities by applying a shear rate up to temperatures of 300°C (according to models, see table). With RT-1 and THERMOCELL package, these systems are compatible with ASTM D3236 and D4402.

Here are the available accessories:



Please find below all item as MK-SV spindle and MB-SV chamber with viscosity range (mPa.s):

Measuring Cylinder		Compatible chamber****		Volume (ml)	Shear rate (s-1)	B-ONE LR	B-ONE
Designation	Item	Designation	Item				
MK-SV414*	116114	MB-SV6R*	116206	3	0,4N	44 to 5,8M	440 to 95M
MK-SV415*	116115	MB-SV7R*	116207	4.4	0,48N	15 to 2M	155 to 33M
MK-SV416*	116116	MB-SV8R*	116208	4.6	0,29N	39 to 5,2M	394 to 85M
MK-SV418	116118	MB-SV13R MB-SV13RC** MB-SVD***	116213 116214 116513	7.5	1,32N	1 to 120K	9 to 1,9M
MK-SV421	116121			8	0,93N	1 to 188K	14 to 3M
MK-SV425	116125			10	0,22N	174 to 23M	1,7K to 377M
MK-SV427	116127			12	0,34N	7 to 0,99M	75 to 16M
MK-SV428	116128			13	0,28N	15 to 1,9M	147 to 31M
MK-SV429	116129			13	0,25N	29 to 3,9M	294 to 63M
MK-SV431	116131			11	0,34N	10 to 1,3M	100 to 21M
MK-SV434	116134			11	0,28N	19 to 2,5M	194 to 41M
MK-SVC	116002			13	0,43N	3 to 420K	32 to 6,8M
MK-SVTR8	140008			8	0,92N	1 to 190K	14 to 3M
MK-SVTR9	140009			12	0,34N	7 to 0,99M	75 to 16M
MK-SVTR10	140010			13	0,28N	15 to 1,9M	146 to 31M
MK-SVTR11	140011			13	0,25N	30 to 3,9M	300 to 64M

Measuring cylinder can be used with different chambers and can be categorized as below:

- Category A: Spindle used with chamber MB-SVD.
- Category B: Spindle used with chamber MB-SV13R, MB-SV6R, MB-SV7R and MB-SV8R.
- Category C: Spindle used with chamber MB-SV13RC.

Measuring systems of category A can only be used with temperature controllers EVA DIN, EVA DIN MS-R, RT-1 PLUS and water jacket CT-LC and CT DIN. They come with disposable cups (a batch of 100) ideal for measuring on aggressive or difficult to clean products.

Measuring systems of category B can only be used with temperature controllers EVA DIN, EVA DIN MS-R, RT-1 PLUS and water jacket CT-LC and CT DIN. They come with a reusable stainless steel measuring cup.

The measuring systems in category C can be used alone or with temperature controllers EVA DIN, EVA DIN MS-R and water jacket CT-LC and CT DIN. Temperature max will be 80°C. They come with a measuring cup in stainless steel and Delrin cap.

### 3.7.1 USE OF MEASURING SYSTEM OF CATEGORY A

Whatever the measurement system model, the using is the same.

If you have temperature unit as RT-1 PLUS or EVA MS-DIN, the measuring head must first be installed on the heating unit (see installation instructions supplied with the temperature control). If you are using water jacket CT-LC or CT-DIN, please place it below instrument installed on its own stand.

The first step is to place the screw on the centring piece (ST-R) and then install the latter on the base of the viscometer.



The screw locks the centring piece on the base of the viscometer.

Make a zero of your viscometer (see section 2.2.4.) if you are using standard model.

Insert the measuring system with the bayonet coupling in the motor shaft (see section 3.1).

Make a zero of your viscometer (see section 2.2.4.) if you are using LR model.





Place the disposable cup holder H-SV13RD in well of the temperature control.

Take a disposable cup MB-SV13RD, fill it with your product (see table section 3.7 for the volume of product to be put into the measurement system). Then place it in the disposable cup holder. Turn it so that the notch at the bottom fits snugly into the bottom of the disposable cup holder.



Use handle and screw of arm (see section 1.2) to gently lower the measuring head. Take care of the Pt100 (if you have one on device) and be sure that it will fit correctly hole of cup.

The centring piece must completely cover the edge of the disposable cup insert.



When the measuring head is properly positioned, there is no need to lock the position using the screw on the arm. Also, be sure to allow sufficient time for stabilization if the measurement temperature is different from the room temperature (at least 10 minutes) before taking your measurement.

You can do your measurement (see section 2.2.2).

Once the measurement is complete, unhook the mobile from the viscometer axis and rise the measuring head. Remove the spindle to clean it.



Remove the holder with the disposable cup by using special tool provide for that operation (provided with Thermocell package). Place holder with disposable cup inside on flat table. Disposable cup will move up from holder. Also clean the temperature probe and the centring piece. It can stay in place on the basis of the viscometer for a next measurement.

### 3.7.2 USE OF MEASURING SYSTEM OF CATEGORY B

Whatever the measurement system model, the using is the same.

If you have temperature unit as RT-1 PLUS or EVA MS-DIN, the measuring head must first be installed on the heating unit (see installation instructions supplied with the temperature control). If you are using water jacket CT-LC or CT-DIN, please place it below instrument installed on its own stand.



The first step is to place the screw on the centring piece (ST-R) and then install the latter on the base of the viscometer.

The screw locks the centring piece on the base of the viscometer.

Make a zero of your viscometer (see section 2.2.4.) if you are using standard model.

Insert the measuring system with the bayonet coupling in the motor shaft (see section 3.1).

Make a zero of your viscometer (see section 2.2.4.) if you are using LR model.



You can then put the product to be measured in the cup MB-SVXR. The required volume is indicated in the table in section 3.7 according to the system used.



Install MB-SVXR chamber on the well of temperature unit or water jacket as shown below.

Use handle and screw of arm (see section 1.2) to gently lower the measuring head. Take care of the Pt100 (if you have one on device) and be sure that it will fit correctly hole of cup.





The centring piece must completely cover the edge of the disposable cup insert.

When the measuring head is properly positioned, there is no need to lock the position using the screw on the arm. Also, be sure to allow sufficient time for stabilization if the measurement temperature is different from the room temperature (at least 10 minutes) before taking your measurement.

You can do your measurement (see section 2.2.2).

When your measurement is complete, it is recommended to remove the cylinder from the viscometer shaft. It will rest in the tube. Rise the measuring head to the highest position by blocking the arm with the screw provided (see section 1.2). Remove the spindle to clean it.




Remove the cup by using special tool provide for that operation (provided with Thermocell). Also clean the temperature probe and the centring piece. It can stay in place on the basis of the viscometer for a next measurement.

### 3.7.3 USE OF MEASURING SYSTEM OF CATEGORY C

These systems can therefore be used with or without a heating unit (RT-1PLUS, DIN EVA and water jacket CT-DIN/CT-LC). The cup MB-SV13RC used for these systems is fixed directly to the base of the viscometer.

Please note that maximum temperature with this system is 80°C.

If you have temperature unit as RT-1 PLUS or EVA MS-DIN, the measuring head must first be installed on the heating unit (see installation instructions supplied with the temperature control). If you are using water jacket CT-LC or CT-DIN, please place it below instrument installed on its own stand.



The first step is to install the Delrin cap on tube MB-SV13RC as shown in the photo below. Also check that the gasket is properly installed on the plug. The first insertion of the cap can be difficult. You must use some silicone grease to facilitate installation.

You can then put the product to be measured in the cup. The necessary volume is indicated in the table in section 3.7 according to the system used.

Make a zero of your viscometer (see section 2.2.4.) if you are using standard model.

Insert the measuring system with the bayonet coupling in the motor shaft (see section 3.1).

Make a zero of your viscometer (see section 2.2.4.) if you are using LR model.



You can then present the tube with the product from the bottom up slowly and making sure to align with the cylinder. Also place the mark on the tube facing you as shown in the picture below). When you are close enough to the Viscometer base, you must rotate the tube to place the pin in the mark of the tube.

Once the tube is properly installed, check that the temperature sensor (if you have one) on the viscometer is at least one centimeter deep into the liquid.

When using your viscometer with a temperature control system (EVA MS DIN or CT DIN), you must then lower the measuring head so that the tube fits into the chamber provided for this purpose. Check that there is no product on the outer wall of the tube.



Use the handle and screw on arm (see section 1.2) to gently lower the measuring head. The tube should be easily housed in the hole of temperature controller. If this is not the case, do not force it and contact your local distributor or LAMY RHEOLOGY.

When the measuring head is properly positioned, there is no need to lock the position using the screw on the stem. Also, be sure to allow sufficient time for stabilization if the measurement temperature is different from the room temperature (at least 10 minutes) before taking your measurement.

Once the setup is complete, you can do your measurement (see section 2.2.2).

When your measurement is complete, it is recommended to remove the cylinder from the viscometer shaft. It will rest in the tube. Rise the measuring head to the highest position by blocking the stem with the screw provided (see section 1.2). Then remove the tube containing the product and the cylinder, take care if the temperature is high. You can then remove the cylinder from the tube to clean it. Remove the cap from the tube to clean it. Clean the temperature sensor.

### 3.8. MS ULV

Measuring system for low viscosities usable with instruments LR version (Stainless steel).

This system makes it possible to measure low viscosity products in control by applying a shear rate. Its advantage is to be compatible with instruments in LR version unlike all other measuring systems.

Here are the available measuring system:

Name	Part number	Dim. (mm)	
MK-C19	116015P	Ø 19	
C Tube with insert	116001	Ø 20	
Delrin cap	116005	-	
C Insert	111934	-	
ST-R centring tool	114436	-	
MB-C Alu Cup	114306	Ø 20	

All this part can be combined to create two different measuring system.

Designation	Part Number	Included					Volume (ml)	Viscosity range (mPa.s)
		Spindle	Cup	Holder	Cap	Tool		
MS-ULV*	116030	116015P	116001		116005		11	1 to 26K
MS-ULV/D**	116031	116015P	114306	111934		114436	11	

K for thousand

\* Not compatible with oven RT1. Can be used without temperature unit.

\*\* Include 100 disposable cup (114306). Must be used with temperature unit or water jacket.


The part number 116030 includes item 116015P, 116001 and 116005. This system can be used with temperature control EVA DIN PLUS, CT-LC and CT DIN or alone without any other accessory.

The part number 116031 is dedicated for measurement with disposable cup and includes item 116015P, 111934, 114436 and 114306 (100 disposable cup). It must be used with a temperature control as EVA DIN, CT DIN or RT-1 PLUS and can't be used alone.

### 3.8.1. USE OF ITEM 116030

This system can be used with or without a heating unit (DIN EVA and CT DIN). The cup used for these systems is the PN116001. This is fixed directly to the basis of the viscometer.

The measuring head must first be installed on the temperature control unit (see installation instructions supplied with the temperature control) or on the standard support if you not use a heating system (see section 1.5).



The first step is to install the cap on the tube as shown in the photo below. Also check that the gasket is properly installed on the plug. The first insertion of the cap can be difficult. You must use some silicone grease to facilitate installation.

You can then put the product to be measured in the cup. The necessary volume is indicated in the table in section 3.8 according to the system used.

Make a zero of your viscometer (see section 2.2.4.) if you are using standard model.

Insert the measuring system with the bayonet coupling in the motor shaft (see section 3.1).

Make a zero of your viscometer (see section 2.2.4.) if you are using LR model.



You can then present the tube with the product from the bottom up slowly and making sure to align with the cylinder. Also place the mark on the tube facing you as shown in the picture below). When you are close enough to the Viscometer base, you must rotate the tube to place the pin in the mark of the tube.

Once the tube is properly installed, check that the temperature sensor (if you have one) on the viscometer is at least one centimetre deep into the liquid.

When using your viscometer with a temperature control system (EVA MS DIN or CT DIN), you must then lower the measuring head so that the tube fits into the chamber provided for this purpose. Check there is no product on the outer wall of the tube.



Use the handle and screw on arm (see section 1.2) to gently lower the measuring head. The tube should be easily housed in the hole of temperature controller. If this is not the case, do not force it and contact your local distributor or LAMY RHEOLOGY. When the measuring head is properly positioned, there is no need to lock the position using the screw on the stem. Also, be sure to allow sufficient time for stabilization if the measurement temperature is different from the room temperature (at least 10 minutes) before taking your measurement

Once the setup is complete, you can do your measurement (see section 2.2.2).

When your measurement is complete, it is recommended to remove the cylinder from the viscometer shaft. It will rest in the tube. Rise the measuring head to the highest position by blocking the stem with the screw provided (see section 1.2). Then remove the tube containing the product and the cylinder, being careful if the temperature is high. You can then remove the cylinder from the tube to clean it. Remove the cap from the tube to clean it. Clean the temperature sensor.

### **3.8.2. USE OF ITEM 116031**

The measuring head must first be installed on the heating unit (see installation instructions supplied with the temperature control).



The first step is to place the screw on the centring piece and then install the latter on the base of the viscometer. The screw locks the centring piece on the basis of the viscometer.

Make a zero of your viscometer (see section 2.2.4.) if you are using standard model.

Insert the measuring system with the bayonet coupling in the motor shaft (see section 3.1).

Make a zero of your viscometer (see section 2.2.4.) if you are using LR model.



Place the disposable cup holder in the well of the temperature control.



Take a disposable cup, fill it with your product (see table on the previous page for the volume of product to be put into the measurement system). Then place it in the disposable cup holder.

Use handle and screw of arm (see section 1.2) to gently lower the measuring head. The centring piece must completely cover the edge of the disposable cup insert.



When the measuring head is properly positioned, this is no required to lock the position using the screw on the arm. Also, be sure to allow sufficient time for stabilization if the measurement temperature is different from the room temperature (at least 10 minutes) before taking your measurement.

You can do your measurement (see section 2.2.2).


Once the measurement is complete, unhook the mobile from the viscometer axis and rise the measuring head. Remove the mobile to clean it. Remove the disposable cup. Also clean the temperature probe and the centring piece. It can stay in place on the basis of the viscometer for a next measurement.

### 3.9. T-BARS AND HELIPRO STAND

The HELIPRO device makes it possible to measure viscosity of gels, pastes, creams and more generally products that do not flowing.

When a measuring mobile turns in this type of product, we can observe appearance of cavities around the rotating spindle which has effect of reducing measured viscosity values. Through its movement vertical, HELIPRO system with its mobile T shape will prevent formation of cavities within product and will ensure reliable and consistent measurement.

This accessory is compatible with viscometers and rheometers from the LAMY RHEOLOGY range except FIRST PRODIG CP 1000, RM 100 CP 1000/2000 PLUS, RM 200 CP 4000 PLUS, GT-300 PLUS, GT-300 PRODIG and DSR 500 CP 4000 PLUS.



THE HELIPRO SYSTEM is delivered with a box including 6 T-bar spindles and 1 adaptation to fix the spindles. It is also delivered with a Hex screw driver for adjust the lower and upper switch limits of the up and down moving, one handle and button already installed on arm.


You can find below measuring for each T-Bars provide with HELIPRO STAND.

T-Bars**	Part number	B-ONE LR (mPa.s)		B-ONE (mPa.s)	
		Minimum	Maximum	Minimum	Maximum
T-A (91)	18091	93	740K	930	12M
T-B (92)	18092	186	1.5M	1,9K	25M
T-C (93)	18093	464	3.7M	4,6K	60M
T-D (94)	18094	1K	7.5M	9,2K	120M
T-E (95)	18095	1,9K	15M	18,6K	240M
T-F (96)	18096	4,6K	37M	46,4K	600M

M for million, K for thousand

\* Viscosity value done for speed range from 0,3 to 15 rpm.

\*\* Included with HELIPRO package (P.N. 111015).



Fixing button

Handle

Install the viscometer on the stand and fix it with the button. Take care of temperature probe if your device gets one.


Insert the spindle inside chuck, lock it by screwing.




Make a zero of your viscometer (see section 2.2.4.) if you are using standard model.

Insert the measuring system with the bayonet coupling in the motor shaft (see section 3.1).

Make a zero of your viscometer (see section 2.2.4.) if you are using LR model.



Place the container with product to be measured below and adjust position of measuring head so that spindle reaches desired position. A release button and handle help you lower the measuring head.



Adjust the position of the 2 limit switch up and down with the screw driver to fix path should be done by measuring head.

Press the button ON/OFF for starting the up and down movement. Helipro stand moves at speed 1 mm/s.

Start the measurement at the desired speed (15 rpm maximum) and choose the right measuring system (see section 2.2.2).

The torque measurement is indicated on the instrument screen using a gauge. Ensure that the measured torque is always sufficiently far from the lower and upper limits (at least 5% above and below). If this is not the case, you can either change the speed.

Then your measurement is finished, stop the movement of Helipro stand, raise the measuring head by using handle and release button and lock it in high position. Remove the measuring spindle to clean it.

#### 4. VERIFICATION OF YOUR DEVICE

Your instrument is calibrated at the factory with an ASTM R2 mobile or MS DIN11 measuring system (see calibration certificate) and a certified oil with a viscosity close to 1000 mPa.s. The verification method differs depending on the measurement system selected. You may decide to perform the verification with your own measurement systems, but it is highly recommended to use one of the two measurement systems mentioned above. In case other systems are used, please contact LAMY RHEOLOGY for the most appropriate verification method.

##### **Viscosity measurement on a 1000 mPa.s standard silicon oil with a ASTM 2555 RV2 measuring system.**

- Fill the 600ml beaker with the standard oil.
- Insert the 600ml beaker in a controlled temperature unit like EVA LR system or thermostatic bath.
- Make a zero of your viscometer (see section 2.2.4.) if you are using standard model.
- Insert the measuring system with the bayonet coupling in the motor shaft (see section 3.1 and 3.2).
- Make a zero of your viscometer (see section 2.2.4.) if you are using LR model.
- Immerge the spindle in the oil at the good level (mark on the spindle, see section 3.2).
- Wait for 15 minutes until the standard oil rises to the good temperature.
- Select on the instrument the measuring system R2, select 50 rpm for the speed, select 60 seconds for the measuring time, and start the measurement (see section 2.2.2).

##### **Viscosity measurement on a 1000 mPa.s standard silicon oil with a defined DIN11 measuring system.**

- Fill the measuring tube DIN 1 with the standard oil.
- Make a zero of your viscometer (see section 2.2.4.) if you are using standard model.
- Insert the measuring system with the bayonet coupling in the motor shaft (see section 3.1).
- Make a zero of your viscometer (see section 2.2.4.) if you are using LR model.
- Fix the tube MB-DIN1 to the viscometer (see section 3.6.1)
- Insert the measuring system in a controlled temperature unit like a EVA DIN system or CT-LC/CT-DIN with thermostatic bath.
- Wait for 10 minutes until the standard oil rises to the good temperature.
- Select on the instrument the measuring system DIN11, select 50s-1 for the speed, select 30 seconds for the measuring time, and start the measurement (see section 2.2.2).

**Result at the end of the measurement must be within +/-5% of the standard viscosity value. If the measure is out, your instrument might need to be recalibrated.**

**Check if the error does not come from a wrong filling, a wrong zero adjustment, a wrong spindle rotation, or a wrong temperature value.**



**LAMY RHEOLOGY**

11 A, rue des Aulnes  
69410 Champagne au Mont d'Or (France)

Tel : 33 (0)4 78 08 54 06

Fax : 33 (0)4 78 08 69 44

[contact@lamyrheology.com](mailto:contact@lamyrheology.com)